



# CERTIFICATE OF CALIBRATION



## *Glanford Electronics Ltd.*

Glanford House,  
Exmoor Avenue  
Skippingdale Business Park,  
Scunthorpe  
North Lincolnshire  
DN15 8NJ

Telephone: (01724) 289600 Fax: (01724) 289008  
Email: [info@glanfordelectronics.com](mailto:info@glanfordelectronics.com) [www.glanfordelectronics.com](http://www.glanfordelectronics.com)

**CALIBRATION:** 41782

*Issued to:* Trace Measurement Limited  
Essex

*Instrument Certified:* ETI Reference Thermometer

*Serial No:* D11120985

*Date in:* 02-Jun-16

*Cal Date:* 20-Jun-16

*Date Due:* 20-Jun-17

*All measurements are traceable to UK National Standards / a UKAS Calibrated Test Instrument*

(GEL178) Jofra DTI-1000 A/B Temperature Indicator S/N:555791-00179, (GEL179) Jofra STS100A901 Temperature Probe S/N:559634-05

*Observations/Corrections*

No Adjustment

*Calibrated at* 22 °C 56 %RH

Page 1 of 2

Certified that, unless otherwise stated, the equipment conforms in all respects to manufacturers published specifications at the points measured  
Calibrations performed in accordance with our registration to BS EN ISO9001:2008  
and the general requirements of BS EN ISO 17025:2005

*certifying engineer* .....

S. Lawrence  
J. Taylor  
M. Gibson ✓  
G. Cowling

FORM 12